

# BIOSAFETY CABINET USE - IMPORTANT NOTES

## • Planning to move the BSC?

- **Contact EHS first!** It must be decontaminated by a professional before it can be moved and re-certified in the new location

## • Has it passed annual certification?

- **Check the certification sticker!** If it failed or missed certification, **DO NOT USE**. Contact EHS!

## • **DO NOT USE** your BSC if:

- It hasn't passed an annual certification
- The sash is broken
- It is hooked up to a flammable gas
- It is not working correctly



**Contact the Biosafety Officer:**  
**404-894-6120**  
**[biosafety@ehs.gatech.edu](mailto:biosafety@ehs.gatech.edu)**

